

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013532.D
 Acq On : 22 Nov 2019 14:49
 Operator : JC/SP
 Sample : K5958-08DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FD-111919DL

Quant Time: Nov 22 15:35:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	684141	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1114323	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1005094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	466531	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	407352	50.26	ug/l	0.00
Spiked Amount			Recovery	=	100.52%	
35) Dibromofluoromethane	5.49	113	334090	47.80	ug/l	0.00
Spiked Amount			Recovery	=	95.60%	
50) Toluene-d8	8.71	98	1340591	50.88	ug/l	0.00
Spiked Amount			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	11.13	95	462343	48.62	ug/l	0.00
Spiked Amount			Recovery	=	97.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
67) Ethyl Benzene	10.25	91	196085	5.583	ug/l	99
68) m/p-Xylenes	10.36	106	75834	5.710	ug/l	100
73) Isopropylbenzene	11.01	105	423272	13.684	ug/l	99
78) n-propylbenzene	11.35	91	921572	26.746	ug/l	100
80) 1,3,5-Trimethylbenzene	11.50	105	582807	22.621	ug/l	100
83) tert-Butylbenzene	11.76	119	52276	2.016	ug/l	86
84) 1,2,4-Trimethylbenzene	11.81	105	3117989	120.640	ug/l	98
85) sec-Butylbenzene	11.94	105	312276	10.476	ug/l	97
86) p-Isopropyltoluene	12.06	119	151457	5.518	ug/l	99
89) n-Butylbenzene	12.39	91	221107	9.232	ug/l #	59
95) Naphthalene	13.83	128	92627	2.881	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013532.D
Acq On : 22 Nov 2019 14:49
Operator : JC/SP
Sample : K5958-08DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FD-111919DL

Quant Time: Nov 22 15:35:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

