

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013540.D
 Acq On : 22 Nov 2019 17:55
 Operator : JC/SP
 Sample : K5978-02 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 14B-2

Quant Time: Nov 25 02:53:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	108623	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	607759	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	523107	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	239558	29.87	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.57%
60) 4-Bromofluorobenzene	11.14	95	229825	27.33	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.10%
63) Toluene-d8	8.71	98	717668	30.34	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.13%

Target Compounds

					Ovalue
25) Chloroform	5.20	83	264766	20.894	ug/l 98
41) Bromodichloromethane	7.89	83	35314	3.734	ug/l 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013540.D
Acq On : 22 Nov 2019 17:55
Operator : JC/SP
Sample : K5978-02 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
14B-2

Quant Time: Nov 25 02:53:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

