

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013557.D
 Acq On : 23 Nov 2019 01:16
 Operator : JC/SP
 Sample : K5962-07DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20BR-20191120DL

Quant Time: Nov 23 03:11:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	650337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1046779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	929049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	423692	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	384249	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
35) Dibromofluoromethane	5.49	113	315184	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	8.71	98	1252887	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
62) 4-Bromofluorobenzene	11.13	95	423010	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
Target Compounds						
4) Vinyl Chloride	1.40	62	4764	0.515	ug/l	94
27) cis-1,2-Dichloroethene	4.58	96	32521	4.470	ug/l	90
44) Trichloroethene	7.21	130	10080	1.309	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013557.D
Acq On : 23 Nov 2019 01:16
Operator : JC/SP
Sample : K5962-07DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20BR-20191120DL

Quant Time: Nov 23 03:11:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

