

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013561.D
 Acq On : 23 Nov 2019 02:49
 Operator : JC/SP
 Sample : K5962-11 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 56BR-20191120

Quant Time: Nov 25 06:17:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	631876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1028746	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	914300	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	421326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	381172	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
35) Dibromofluoromethane	5.49	113	316428	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	8.71	98	1239679	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.13	95	417540	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
Target Compounds						
4) Vinyl Chloride	1.40	62	7645	0.851	ug/l #	85
27) cis-1,2-Dichloroethene	4.59	96	193182	27.326	ug/l	90
44) Trichloroethene	7.21	130	1366412	180.569	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013561.D
Acq On : 23 Nov 2019 02:49
Operator : JC/SP
Sample : K5962-11 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
56BR-20191120

Quant Time: Nov 25 06:17:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

