

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013566.D
 Acq On : 23 Nov 2019 04:46
 Operator : JC/SP
 Sample : K5968-06 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-118.5N-20191121

Quant Time: Nov 25 06:48:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	636410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1029934	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	920047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	414942	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	385724	51.16	ug/l	0.00
Spiked Amount			Recovery	=	102.32%	
35) Dibromofluoromethane	5.49	113	311995	48.30	ug/l	0.00
Spiked Amount			Recovery	=	96.60%	
50) Toluene-d8	8.71	98	1238764	50.87	ug/l	0.00
Spiked Amount			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.14	95	410962	46.76	ug/l	0.00
Spiked Amount			Recovery	=	93.52%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013566.D
Acq On : 23 Nov 2019 04:46
Operator : JC/SP
Sample : K5968-06 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-118.5N-20191121

Quant Time: Nov 25 06:48:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

