

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013573.D
 Acq On : 23 Nov 2019 07:29
 Operator : JC/SP
 Sample : K5968-13 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INTERCEPTOR-20191121

Quant Time: Nov 25 07:15:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	613800	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1004507	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	888279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	397115	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	370714	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
35) Dibromofluoromethane	5.48	113	304437	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
50) Toluene-d8	8.71	98	1203493	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
62) 4-Bromofluorobenzene	11.14	95	394208	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
Target Compounds						
11) Tert butyl alcohol	3.01	59	11775	6.144	ug/l #	82
27) cis-1,2-Dichloroethene	4.58	96	4616	0.672	ug/l	77
44) Trichloroethene	7.22	130	8610	1.165	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013573.D
Acq On : 23 Nov 2019 07:29
Operator : JC/SP
Sample : K5968-13 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INTERCEPTOR-20191121

Quant Time: Nov 25 07:15:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

