

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013522.D
 Acq On : 22 Nov 2019 10:45
 Operator : JC/SP
 Sample : VX1122WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1122WBL01

Quant Time: Nov 22 14:19:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.00	128	100173	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	562651	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	470760	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.04	65	212081	28.67	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	95.57%
60) 4-Bromofluorobenzene	11.14	95	216225	28.58	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	95.27%
63) Toluene-d8	8.71	98	647280	30.41	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.37%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013522.D
Acq On : 22 Nov 2019 10:45
Operator : JC/SP
Sample : VX1122WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1122WBL01

Quant Time: Nov 22 14:19:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 30.000 ug/l

RT: 5.00 min Scan# 674

Delta R.T. -0.01 min

Lab File: VX013522.D

Acq: 22 Nov 2019 10:45

Instrument :

MSVOA_X

ClientSampleId :

VX1122WBL01


Tot Ion:128 Resp: 100173

Ion Ratio Lower Upper

128 100

49 152.5 0.0 388.3




130 124.2 0.0 318.5



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 49.00 (48.70 to 49.70): VX0

Ion 130.00 (129.70 to 130.70): V



#27

1,2-Dichloroethane-d4

Concen: 28.671 ug/l

RT: 6.04 min Scan# 846

Delta R.T. -0.01 min

Lab File: VX013522.D


Acq: 22 Nov 2019 10:45

Tgt Ion: 65 Resp: 212081

Ion Ratio Lower Upper



65 100


67 53.3 41.0 61.6



Abundance Ion 65.00 (64.70 to 65.70): VX0

Ion 67.00 (66.70 to 67.70): VX0









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 6.84 min Scan# 977
Delta R.T. -0.01 min
Lab File: VX013522.D
Acq: 22 Nov 2019 10:45




Instrument :
MSVOA_X
ClientSampleId :
VX1122WBL01


Tot Ion	Ratio	Lower	Upper
114	100		
63	19.6	16.1	24.1
88	15.4	12.3	18.5



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.10 min Scan# 1512
Delta R.T. -0.01 min
Lab File: VX013522.D
Acq: 22 Nov 2019 10:45

Tgt Ion	Ratio	Lower	Upper
117	100		
82	54.6	43.7	65.5
119	32.1	25.5	38.3









#60
4-Bromofluorobenzene
Concen: 28.577 ug/l
RT: 11.14 min Scan# 1681
Delta R.T. -0.00 min
Lab File: VX013522.D
Acq: 22 Nov 2019 10:45

Instrument :
MSVOA_X
ClientSampleId :
VX1122WBL01

Tot Ion	Ratio	Lower	Upper
95	100		
174	90.1	70.2	105.2
176	88.2	68.3	102.5



#63
Toluene-d8
Concen: 30.406 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. -0.00 min
Lab File: VX013522.D
Acq: 22 Nov 2019 10:45

Tgt Ion	Ratio	Lower	Upper
98	100		
100	65.4	52.5	78.7

