

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032981.D
 Acq On : 22 Nov 2022 15:58
 Operator : JC/MD
 Sample : VX1122MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1122MBL01

Quant Time: Nov 23 05:49:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189391	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79756	68.453	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 136.900%#	
35) Dibromofluoromethane	5.385	113	62586	58.617	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.240%	
50) Toluene-d8	8.647	98	214452	67.510	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 135.020%#	
62) 4-Bromofluorobenzene	11.079	95	76396	54.469	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.940%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX032981.D
Acq On : 22 Nov 2022 15:58
Operator : JC/MD
Sample : VX1122MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1122MBL01

Quant Time: Nov 23 05:49:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

