

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032988.D
 Acq On : 22 Nov 2022 19:09
 Operator : JC/MD
 Sample : N5729-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20221117DL

Quant Time: Nov 23 05:50:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163376	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38582	32.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	65.880%#
35) Dibromofluoromethane	5.391	113	41890	39.524	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	79.040%
50) Toluene-d8	8.647	98	86903	26.225	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	52.460%#
62) 4-Bromofluorobenzene	11.079	95	63716	45.765	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.520%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1886	1.734	ug/l	98
12) 1,1-Dichloroethene	2.325	96	704	0.695	ug/l	94
44) Trichloroethene	7.129	130	39692	29.769	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX032988.D
Acq On : 22 Nov 2022 19:09
Operator : JC/MD
Sample : N5729-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20221117DL

Quant Time: Nov 23 05:50:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

