

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032989.D
 Acq On : 22 Nov 2022 19:32
 Operator : JC/MD
 Sample : N5729-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20221117DL

Quant Time: Nov 23 05:50:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36799	31.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	62.520%#
35) Dibromofluoromethane	5.385	113	40260	37.729	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	75.460%
50) Toluene-d8	8.647	98	83593	24.955	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	49.920%#
62) 4-Bromofluorobenzene	11.079	95	62003	44.233	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.460%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	2410	2.216	ug/l	95
12) 1,1-Dichloroethene	2.325	96	376	0.371	ug/l	94
44) Trichloroethene	7.129	130	45434	33.846	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 23 05:50:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

