

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032991.D
 Acq On : 22 Nov 2022 20:18
 Operator : JC/MD
 Sample : N5729-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20221117

Quant Time: Nov 23 05:50:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	182810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36558	31.710	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	63.420%#
35) Dibromofluoromethane	5.385	113	40071	38.881	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	77.760%
50) Toluene-d8	8.647	98	82902	25.685	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	51.380%#
62) 4-Bromofluorobenzene	11.079	95	61319	45.293	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.580%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	548	0.630	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.337	101	6084	5.703	ug/l	97
12) 1,1-Dichloroethene	2.319	96	737	0.742	ug/l #	86
27) cis-1,2-Dichloroethene	4.501	96	1120	0.867	ug/l	97
32) 1,1,1-Trichloroethane	5.391	97	932	0.465	ug/l #	49
44) Trichloroethene	7.129	130	87913	67.807	ug/l	100
52) Toluene	8.720	92	2087	0.687	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX032991.D
Acq On : 22 Nov 2022 20:18
Operator : JC/MD
Sample : N5729-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20221117

Quant Time: Nov 23 05:50:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

