

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032992.D
 Acq On : 22 Nov 2022 20:41
 Operator : JC/MD
 Sample : N5729-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT161S-20221117RE

Quant Time: Nov 23 05:50:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	179365	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72446	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35864	31.461	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	62.920%#	
35) Dibromofluoromethane	5.391	113	38375	37.950	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	75.900%	
50) Toluene-d8	8.653	98	83072	26.280	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	52.560%#	
62) 4-Bromofluorobenzene	11.079	95	61478	46.283	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	92.560%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	653	0.665	ug/l	93
19) Methyl tert-butyl Ether	3.111	73	1387	0.407	ug/l	95
24) 1,1-Dichloroethane	3.617	63	5936	3.005	ug/l	98
30) Chloroform	5.111	83	719	0.325	ug/l #	71
44) Trichloroethene	7.135	130	1600	1.258	ug/l	91
64) Tetrachloroethene	9.275	164	372	0.326	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX032992.D
Acq On : 22 Nov 2022 20:41
Operator : JC/MD
Sample : N5729-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT161S-20221117RE

Quant Time: Nov 23 05:50:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

