

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032993.D
 Acq On : 22 Nov 2022 21:04
 Operator : JC/MD
 Sample : N5730-08 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-20221117

Quant Time: Nov 23 05:51:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187946	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57301	50.877	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.760%		
35) Dibromofluoromethane	5.391	113	48270	45.556	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 91.120%		
50) Toluene-d8	8.647	98	145854	45.559	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 91.120%#		
62) 4-Bromofluorobenzene	11.079	95	64463	46.314	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.620%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1730	1.600	ug/l	92
12) 1,1-Dichloroethene	2.319	96	343	0.341	ug/l	92
27) cis-1,2-Dichloroethene	4.507	96	375	0.287	ug/l	56
44) Trichloroethene	7.129	130	132972	99.759	ug/l	97
64) Tetrachloroethene	9.275	164	490	0.418	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX032993.D
Acq On : 22 Nov 2022 21:04
Operator : JC/MD
Sample : N5730-08 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-20221117

Quant Time: Nov 23 05:51:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

