

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX033001.D
 Acq On : 23 Nov 2022 00:08
 Operator : JC/MD
 Sample : N5730-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20221117

Quant Time: Nov 23 05:52:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116636	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187110	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74495	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36136	30.483	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	60.960%#
35) Dibromofluoromethane	5.391	113	38516	36.513	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	73.020%#
50) Toluene-d8	8.647	98	83122	25.116	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	50.240%#
62) 4-Bromofluorobenzene	11.079	95	68158	49.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.380%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	8507	7.786	ug/l	98
12) 1,1-Dichloroethene	2.325	96	1585	1.557	ug/l	88
27) cis-1,2-Dichloroethene	4.489	96	2109	1.594	ug/l	83
30) Chloroform	5.092	83	1031	0.450	ug/l	95
38) Carbon Tetrachloride	5.678	117	726	0.388	ug/l #	73
44) Trichloroethene	7.129	130	1039410	783.274	ug/l	97
64) Tetrachloroethene	9.275	164	6727	5.765	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX033001.D
Acq On : 23 Nov 2022 00:08
Operator : JC/MD
Sample : N5730-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20221117

Quant Time: Nov 23 05:52:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

