

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX033002.D
 Acq On : 23 Nov 2022 00:31
 Operator : JC/MD
 Sample : N5741-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Nov 23 05:52:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189887	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36501	30.537	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	61.080%#
35) Dibromofluoromethane	5.385	113	39677	37.064	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	74.120%#
50) Toluene-d8	8.647	98	83638	24.883	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	49.760%#
62) 4-Bromofluorobenzene	11.079	95	69234	49.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.460%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1124	2.101	ug/l #	76
19) Methyl tert-butyl Ether	3.117	73	24248	6.811	ug/l	100
44) Trichloroethene	7.135	130	3961	2.941	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX033002.D
Acq On : 23 Nov 2022 00:31
Operator : JC/MD
Sample : N5741-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Nov 23 05:52:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

