

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX033003.D
 Acq On : 23 Nov 2022 00:54
 Operator : JC/MD
 Sample : N5741-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Nov 23 05:52:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	185217	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	34007	29.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	58.400%#
35) Dibromofluoromethane	5.385	113	37286	35.708	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	71.420%#
50) Toluene-d8	8.653	98	77240	23.439	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	46.880%#
62) 4-Bromofluorobenzene	11.079	95	65806	47.976	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.960%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1226	2.362	ug/l #	81
44) Trichloroethene	7.129	130	1574	1.198	ug/l	94
86) p-Isopropyltoluene	12.012	119	1566	0.337	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX033003.D
Acq On : 23 Nov 2022 00:54
Operator : JC/MD
Sample : N5741-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Nov 23 05:52:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

