

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX033004.D
 Acq On : 23 Nov 2022 01:17
 Operator : JC/MD
 Sample : N5741-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16R

Quant Time: Nov 23 05:52:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116353	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187305	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76470	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35456	29.928	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	59.860%#
35) Dibromofluoromethane	5.385	113	38370	36.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	72.680%#
50) Toluene-d8	8.653	98	79965	24.049	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	48.100%#
62) 4-Bromofluorobenzene	11.079	95	67983	49.011	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.020%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2354	4.448	ug/l	95
19) Methyl tert-butyl Ether	3.111	73	21801	6.191	ug/l	97
40) Benzene	6.044	78	5849	1.236	ug/l	96
44) Trichloroethene	7.129	130	1014	0.763	ug/l	91
52) Toluene	8.720	92	740	0.238	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX033004.D
Acq On : 23 Nov 2022 01:17
Operator : JC/MD
Sample : N5741-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16R

Quant Time: Nov 23 05:52:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

