

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX033005.D
 Acq On : 23 Nov 2022 01:40
 Operator : JC/MD
 Sample : N5741-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18R

Quant Time: Nov 23 05:52:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	186829	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35658	30.845	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	61.700%#
35) Dibromofluoromethane	5.385	113	38820	36.857	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	73.720%#
50) Toluene-d8	8.647	98	81941	24.767	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	49.540%#
62) 4-Bromofluorobenzene	11.079	95	68711	49.662	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.320%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1408	2.719	ug/l #	85
17) Carbon Disulfide	2.514	76	1299	0.473	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	3795	1.101	ug/l	98
44) Trichloroethene	7.129	130	695	0.525	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX033005.D
Acq On : 23 Nov 2022 01:40
Operator : JC/MD
Sample : N5741-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18R

Quant Time: Nov 23 05:52:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

