

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032980.D
 Acq On : 22 Nov 2022 15:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 05:48:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.382	0.347	9.2	87	0.00
3 P	Chloromethane	0.412	0.387	6.1	91	0.00
4 C	Vinyl Chloride	0.459	0.421	8.3#	87	0.00
5 T	Bromomethane	0.340	0.323	5.0	93	0.01
6 T	Chloroethane	0.374	0.372	0.5	90	0.00
7 T	Trichlorofluoromethane	0.891	0.856	3.9	90	0.00
8 T	Diethyl Ether	0.281	0.298	-6.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.452	3.4	89	0.00
10 T	Methyl Iodide	0.536	0.551	-2.8	88	0.00
11 T	Tert butyl alcohol	0.077	0.092	-19.5	110	0.05
12 CM	1,1-Dichloroethene	0.436	0.425	2.5#	89	0.00
13 T	Acrolein	0.089	0.092	-3.4	100	0.00
14 T	Allyl chloride	0.704	0.699	0.7	90	0.00
15 T	Acrylonitrile	0.218	0.244	-11.9	104	0.00
16 T	Acetone	0.227	0.227	0.0	92	0.01
17 T	Carbon Disulfide	1.205	1.083	10.1	86	0.00
18 T	Methyl Acetate	0.629	0.699	-11.1	105	0.00
19 T	Methyl tert-butyl Ether	1.513	1.633	-7.9	98	0.00
20 T	Methylene Chloride	0.548	0.506	7.7	95	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.485	2.8	91	0.00
22 T	Diisopropyl ether	1.534	1.666	-8.6	95	0.00
23 T	Vinyl Acetate	1.201	1.346	-12.1	98	0.00
24 P	1,1-Dichloroethane	0.878	0.892	-1.6	94	0.00
25 T	2-Butanone	0.314	0.352	-12.1	103	0.00
26 T	2,2-Dichloropropane	0.706	0.708	-0.3	89	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.580	-2.3	91	0.00
28 T	Bromochloromethane	0.386	0.397	-2.8	93	0.00
29 T	Tetrahydrofuran	0.199	0.230	-15.6	107	0.00
30 C	Chloroform	0.983	1.018	-3.6#	93	0.00
31 T	Cyclohexane	0.789	0.771	2.3	88	0.00
32 T	1,1,1-Trichloroethane	0.880	0.887	-0.8	89	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.529	1.9	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.282	0.273	3.2	95	0.00
36 T	1,1-Dichloropropene	0.451	0.437	3.1	89	0.00
37 T	Ethyl Acetate	0.393	0.445	-13.2	103	0.00
38 T	Carbon Tetrachloride	0.500	0.485	3.0	88	0.00
39 T	Methylcyclohexane	0.514	0.509	1.0	90	0.00
40 TM	Benzene	1.264	1.278	-1.1	92	0.00
41 T	Methacrylonitrile	0.221	0.251	-13.6	103	0.00
42 TM	1,2-Dichloroethane	0.509	0.544	-6.9	97	0.00
43 T	Isopropyl Acetate	0.650	0.728	-12.0	101	0.00
44 TM	Trichloroethene	0.355	0.350	1.4	89	0.00
45 C	1,2-Dichloropropane	0.312	0.331	-6.1#	95	0.00
46 T	Dibromomethane	0.244	0.256	-4.9	97	0.00
47 T	Bromodichloromethane	0.471	0.502	-6.6	95	0.00
48 T	Methyl methacrylate	0.335	0.382	-14.0	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032980.D
 Acq On : 22 Nov 2022 15:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 05:48:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.008	-33.3#	120	0.05
50 S	Toluene-d8	0.915	0.866	5.4	97	0.00
51 T	4-Methyl-2-Pentanone	0.401	0.469	-17.0	105	0.00
52 CM	Toluene	0.831	0.849	-2.2#	92	0.00
53 T	t-1,3-Dichloropropene	0.462	0.517	-11.9	93	0.00
54 T	cis-1,3-Dichloropropene	0.508	0.555	-9.3	95	0.00
55 T	1,1,2-Trichloroethane	0.338	0.364	-7.7	99	0.00
56 T	Ethyl methacrylate	0.466	0.534	-14.6	98	0.00
57 T	1,3-Dichloropropane	0.546	0.590	-8.1	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.246	-11.3	100	0.00
59 T	2-Hexanone	0.296	0.353	-19.3	104	0.00
60 T	Dibromochloromethane	0.364	0.398	-9.3	94	0.00
61 T	1,2-Dibromoethane	0.357	0.393	-10.1	98	0.00
62 S	4-Bromofluorobenzene	0.370	0.383	-3.5	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.354	0.333	5.9	91	0.00
65 PM	Chlorobenzene	0.997	0.993	0.4	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.382	-4.7	95	0.00
67 C	Ethyl Benzene	1.768	1.809	-2.3#	92	0.00
68 T	m/p-Xylenes	0.692	0.706	-2.0	93	0.00
69 T	o-Xylene	0.670	0.691	-3.1	93	0.00
70 T	Styrene	1.096	1.177	-7.4	93	0.00
71 P	Bromoform	0.280	0.309	-10.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.573	3.518	1.5	93	0.00
74 T	N-amyl acetate	1.215	1.368	-12.6	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.078	1.121	-4.0	100	0.00
76 T	1,2,3-Trichloropropane	0.947	1.037	-9.5	104	0.00
77 T	Bromobenzene	0.879	0.860	2.2	93	0.00
78 T	n-propylbenzene	4.085	4.101	-0.4	93	0.00
79 T	2-Chlorotoluene	2.481	2.447	1.4	94	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.061	-2.0	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.315	-8.2	99	0.00
82 T	4-Chlorotoluene	2.908	2.907	0.0	94	0.00
83 T	tert-Butylbenzene	2.988	2.975	0.4	92	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.048	-1.8	93	0.00
85 T	sec-Butylbenzene	3.574	3.619	-1.3	92	0.00
86 T	p-Isopropyltoluene	3.028	3.073	-1.5	92	0.00
87 T	1,3-Dichlorobenzene	1.648	1.635	0.8	95	0.00
88 T	1,4-Dichlorobenzene	1.688	1.642	2.7	92	0.00
89 T	n-Butylbenzene	2.609	2.654	-1.7	92	0.00
90 T	Hexachloroethane	0.498	0.500	-0.4	94	0.00
91 T	1,2-Dichlorobenzene	1.603	1.604	-0.1	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.235	-9.3	99	0.00
93 T	1,2,4-Trichlorobenzene	0.988	0.982	0.6	93	0.00
94 T	Hexachlorobutadiene	0.436	0.399	8.5	91	0.00
95 T	Naphthalene	3.197	3.342	-4.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
Data File : VX032980.D
Acq On : 22 Nov 2022 15:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 23 05:48:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.991	0.974	1.7	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6