



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112222\
 Data File : VX032967.D
 Acq On : 22 Nov 2022 08:59
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Title : SW846 8260
 Last Update : Tue Nov 22 13:58:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	6226	PASS
75	95	30	60	55.1	16544	PASS
95	95	100	100	100.0	30008	PASS
96	95	5	9	7.0	2097	PASS
173	174	0.00	2	1.2	282	PASS
174	95	50	100	78.8	23650	PASS
175	174	5	9	7.7	1820	PASS
176	174	95	101	96.8	22894	PASS
177	176	5	9	7.0	1595	PASS