

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112223\
 Data File : VX039005.D
 Acq On : 22 Nov 2023 16:09
 Operator : JC/MD
 Sample : 05407-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01-20231114

Quant Time: Nov 23 01:34:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	91805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77885	52.650	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.300%
35) Dibromofluoromethane	5.391	113	60915	45.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.020%
50) Toluene-d8	8.647	98	242773	46.882	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.760%
62) 4-Bromofluorobenzene	11.079	95	101805	49.112	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.220%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	6564	8.335	ug/l	97
20) Methylene Chloride	2.788	84	2224	1.567	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112223\
Data File : VX039005.D
Acq On : 22 Nov 2023 16:09
Operator : JC/MD
Sample : 05407-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01-20231114

Quant Time: Nov 23 01:34:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

