

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112223\
 Data File : VX039007.D
 Acq On : 22 Nov 2023 16:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Nov 23 01:35:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	90968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78247	53.381	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.760%
35) Dibromofluoromethane	5.385	113	61265	44.781	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.560%
50) Toluene-d8	8.647	98	240464	45.939	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.880%#
62) 4-Bromofluorobenzene	11.079	95	102265	48.805	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.600%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.788	84	1599	1.137	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112223\
Data File : VX039007.D
Acq On : 22 Nov 2023 16:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Nov 23 01:35:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

