

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043951.D
 Acq On : 22 Nov 2024 11:50
 Operator : JC/MD
 Sample : VX1122MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1122MBL01

Quant Time: Nov 22 22:18:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	110756	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214146	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	189507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	94492	49.742	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.480%
35) Dibromofluoromethane	5.379	113	67961	44.414	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.820%
50) Toluene-d8	8.647	98	261229	49.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.060%
62) 4-Bromofluorobenzene	11.079	95	83330	46.420	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.840%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043951.D
Acq On : 22 Nov 2024 11:50
Operator : JC/MD
Sample : VX1122MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1122MBL01

Quant Time: Nov 22 22:18:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

