

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043965.D
 Acq On : 22 Nov 2024 17:16
 Operator : JC/MD
 Sample : P4846-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-EVAL-02D-20241113

Quant Time: Nov 22 22:26:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	106324	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	206768	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69943	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95596	52.420	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.840%
35) Dibromofluoromethane	5.379	113	67360	45.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.180%
50) Toluene-d8	8.647	98	251499	49.382	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.760%
62) 4-Bromofluorobenzene	11.079	95	79415	45.818	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.640%
Target Compounds						
					Qvalue	
3) Chloromethane	1.295	50	556	0.392	ug/l #	69
16) Acetone	2.380	43	1325	1.582	ug/l	100
44) Trichloroethene	7.123	130	2502	1.541	ug/l	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043965.D
Acq On : 22 Nov 2024 17:16
Operator : JC/MD
Sample : P4846-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-02D-20241113

Quant Time: Nov 22 22:26:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

