

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043968.D
 Acq On : 22 Nov 2024 18:25
 Operator : JC/MD
 Sample : P4846-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-EVAL-01D-20241113

Quant Time: Nov 22 22:28:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	108343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	213438	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	185232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97377	52.402	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.800%	
35) Dibromofluoromethane	5.379	113	69840	45.793	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	91.580%	
50) Toluene-d8	8.647	98	260832	49.614	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.220%	
62) 4-Bromofluorobenzene	11.079	95	87967	49.166	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.340%	
Target Compounds						
12) 1,1-Dichloroethene	2.313	96	850	0.681	ug/l	89
16) Acetone	2.374	43	1472	1.725	ug/l #	85
27) cis-1,2-Dichloroethene	4.483	96	14439	9.063	ug/l	88
44) Trichloroethene	7.117	130	571405	340.966	ug/l	97
52) Toluene	8.720	92	1801	0.508	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043968.D
Acq On : 22 Nov 2024 18:25
Operator : JC/MD
Sample : P4846-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-01D-20241113

Quant Time: Nov 22 22:28:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

