

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043969.D
 Acq On : 22 Nov 2024 18:47
 Operator : JC/MD
 Sample : P4846-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-EVAL-01S-20241113

Quant Time: Nov 22 22:28:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	106735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	209074	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	95772	52.315	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.620%
35) Dibromofluoromethane	5.379	113	68870	46.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.200%
50) Toluene-d8	8.647	98	257950	50.090	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.180%
62) 4-Bromofluorobenzene	11.079	95	85223	48.626	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.260%
Target Compounds						
16) Acetone	2.379	43	1230	1.463	ug/l #	88
44) Trichloroethene	7.116	130	147401	89.792	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043969.D
Acq On : 22 Nov 2024 18:47
Operator : JC/MD
Sample : P4846-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-01S-20241113

Quant Time: Nov 22 22:28:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

