

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043971.D
 Acq On : 22 Nov 2024 19:33
 Operator : JC/MD
 Sample : P4873-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-RB-4-20241114

Quant Time: Nov 22 22:29:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105197	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	204646	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	177710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94351	52.292	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.580%
35) Dibromofluoromethane	5.379	113	67160	45.928	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.860%
50) Toluene-d8	8.647	98	252075	50.008	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.020%
62) 4-Bromofluorobenzene	11.079	95	83106	48.445	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.880%
Target Compounds						
16) Acetone	2.380	43	8243	9.947	ug/l #	88
20) Methylene Chloride	2.782	84	587	0.417	ug/l #	77
44) Trichloroethene	7.129	130	1494	0.930	ug/l	94
52) Toluene	8.720	92	1367	0.402	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043971.D
Acq On : 22 Nov 2024 19:33
Operator : JC/MD
Sample : P4873-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-RB-4-20241114

Quant Time: Nov 22 22:29:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

