

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043972.D
 Acq On : 22 Nov 2024 19:56
 Operator : JC/MD
 Sample : P4873-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2

Quant Time: Nov 22 22:30:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	112888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97617	50.416	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.840%		
35) Dibromofluoromethane	5.379	113	69712	45.117	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 90.240%		
50) Toluene-d8	8.647	98	256585	48.174	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 96.340%		
62) 4-Bromofluorobenzene	11.079	95	86524	47.733	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 95.460%		
Target Compounds						
16) Acetone	2.379	43	1465	1.647	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043972.D
Acq On : 22 Nov 2024 19:56
Operator : JC/MD
Sample : P4873-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2

Quant Time: Nov 22 22:30:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

