

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019545.D
 Acq On : 23 Nov 2020 13:08
 Operator : JC/MD
 Sample : L4857-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 24 04:01:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	326561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	522841	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	450955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195105	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	191566	45.24	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.48%
35) Dibromofluoromethane	5.46	113	153564	46.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.58%
50) Toluene-d8	8.70	98	608366	47.70	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.40%
62) 4-Bromofluorobenzene	11.12	95	210636	45.91	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.82%

Target Compounds

					Qvalue
20) Methylene Chloride	2.83	84	2307	0.564 ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019545.D
Acq On : 23 Nov 2020 13:08
Operator : JC/MD
Sample : L4857-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 24 04:01:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

