

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019548.D
 Acq On : 23 Nov 2020 14:18
 Operator : JC/MD
 Sample : L4836-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-01

Quant Time: Nov 24 11:44:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	314586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	508396	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	434716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	190581	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	186696	45.77	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.54%	
35) Dibromofluoromethane		5.46	113	149960	46.99	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.98%	
50) Toluene-d8		8.70	98	588429	47.45	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.90%	
62) 4-Bromofluorobenzene		11.12	95	206682	46.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	8214	5.207	ug/l	99
44) Trichloroethene	7.19	130	1216	0.317	ug/l	81
52) Toluene	8.76	92	3049	0.340	ug/l	95
84) 1,2,4-Trimethylbenzene	11.79	105	4720	0.422	ug/l	98
95) Naphthalene	13.82	128	1437	2.276	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019548.D
Acq On : 23 Nov 2020 14:18
Operator : JC/MD
Sample : L4836-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-01

Quant Time: Nov 24 11:44:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

