

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019550.D
 Acq On : 23 Nov 2020 15:05
 Operator : JC/MD
 Sample : L4798-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 29BR-20201118

Manual Integrations
 APPROVED

MMDadoda
 11/24/2020 2:59:43 PM

Quant Time: Nov 24 05:09:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	318842	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	517445	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	442672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195108	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	188419	45.58	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.16%		
35) Dibromofluoromethane	5.46	113	151545	46.66	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.32%		
50) Toluene-d8	8.70	98	587110	46.52	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	93.04%		
62) 4-Bromofluorobenzene	11.12	95	209488	46.13	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.26%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.18	85	109518m	35.322	ug/l	
4) Vinyl Chloride	1.39	62	30620	7.995	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	2.36	101	93494	30.315	ug/l	99
12) 1,1-Dichloroethene	2.34	96	56326	17.794	ug/l	100
21) trans-1,2-Dichloroethene	3.14	96	41858	11.671	ug/l	97
27) cis-1,2-Dichloroethene	4.56	96	1248850	303.143	ug/l	87
44) Trichloroethene	7.20	130	11436661	2929.351	ug/l	94
64) Tetrachloroethene	9.32	164	1129	0.324	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019550.D
Acq On : 23 Nov 2020 15:05
Operator : JC/MD
Sample : L4798-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
29BR-20201118

Manual Integrations
APPROVED

MMDadoda
11/24/2020 2:59:43 PM

Quant Time: Nov 24 05:09:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

