

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019554.D
 Acq On : 23 Nov 2020 16:39
 Operator : JC/MD
 Sample : L4798-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20201118

Manual Integrations
 APPROVED

MMDadoda
 11/24/2020 2:59:47 PM

Quant Time: Nov 24 05:10:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	301219	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	490239	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	414615	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	175812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	180397	46.19	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.38%		
35) Dibromofluoromethane	5.46	113	144308	46.90	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.80%		
50) Toluene-d8	8.70	98	565105	47.26	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.52%		
62) 4-Bromofluorobenzene	11.12	95	191171	44.44	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.88%		
Target Compounds						
2) Dichlorodifluoromethane	1.18	85	10611m	3.622	ug/l	
4) Vinyl Chloride	1.39	62	302293	83.544	ug/l	98
12) 1,1-Dichloroethene	2.34	96	4632	1.549	ug/l	91
21) trans-1,2-Dichloroethene	3.14	96	4355	1.285	ug/l	98
27) cis-1,2-Dichloroethene	4.56	96	1445498	371.405	ug/l	88
29) Tetrahydrofuran	5.11	42	870	0.601	ug/l #	80
44) Trichloroethene	7.18	130	132778	35.897	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019554.D
Acq On : 23 Nov 2020 16:39
Operator : JC/MD
Sample : L4798-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20201118

Manual Integrations
APPROVED

MMDadoda
11/24/2020 2:59:47 PM

Quant Time: Nov 24 05:10:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

