

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019560.D
 Acq On : 23 Nov 2020 18:59
 Operator : JC/MD
 Sample : L4839-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B4-GW

Quant Time: Nov 24 04:39:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	320692	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	516212	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	438708	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	173425	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	189209	45.50	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.00%	
35) Dibromofluoromethane		5.46	113	151482	46.75	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.50%	
50) Toluene-d8		8.70	98	600904	47.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.44%	
62) 4-Bromofluorobenzene		11.12	95	195946	43.25	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	7471	4.646	ug/l	98
20) Methylene Chloride	2.83	84	1631	0.406	ug/l	86
27) cis-1,2-Dichloroethene	4.56	96	6012	1.451	ug/l	74
44) Trichloroethene	7.18	130	13049	3.350	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019560.D
Acq On : 23 Nov 2020 18:59
Operator : JC/MD
Sample : L4839-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B4-GW

Quant Time: Nov 24 04:39:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

