

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019561.D
 Acq On : 23 Nov 2020 19:23
 Operator : JC/MD
 Sample : L4839-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B5-GW

Manual Integrations
 APPROVED

MMDadoda
 11/24/2020 2:59:54 PM

Quant Time: Nov 24 12:00:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	312651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	503114	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	417361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	174046	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	187640	46.29	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.58%		
35) Dibromofluoromethane	5.46	113	150130	47.54	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.08%		
50) Toluene-d8	8.70	98	575375	46.89	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	93.78%		
62) 4-Bromofluorobenzene	11.12	95	191154	43.29	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	86.58%		
Target Compounds						
16) Acetone	2.43	43	4435	2.829	ug/l	91
18) Methyl Acetate	2.75	43	1099m	0.917	ug/l	
20) Methylene Chloride	2.82	84	1171	0.299	ug/l #	88
27) cis-1,2-Dichloroethene	4.56	96	4535	1.123	ug/l	56
44) Trichloroethene	7.18	130	10012	2.637	ug/l	99
95) Naphthalene	13.82	128	1767	2.310	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019561.D
Acq On : 23 Nov 2020 19:23
Operator : JC/MD
Sample : L4839-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
B5-GW

Manual Integrations
APPROVED

MMDadoda
11/24/2020 2:59:54 PM

Quant Time: Nov 24 12:00:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

