

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019567.D
 Acq On : 23 Nov 2020 22:29
 Operator : JC/MD
 Sample : VX1123WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1123WBL02

Quant Time: Nov 24 05:04:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	329916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	534741	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	468888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	201198	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	198906	46.50	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.00%
35) Dibromofluoromethane	5.46	113	155926	46.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.90%
50) Toluene-d8	8.70	98	631272	48.40	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.80%
62) 4-Bromofluorobenzene	11.12	95	215720	45.97	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.94%

Target Compounds					Qvalue
44) Trichloroethene	7.18	130	13471	3.339 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019567.D
Acq On : 23 Nov 2020 22:29
Operator : JC/MD
Sample : VX1123WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1123WBL02

Quant Time: Nov 24 05:04:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

