

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019570.D
 Acq On : 23 Nov 2020 23:39
 Operator : JC/MD
 Sample : PB133157TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157TB

Quant Time: Nov 24 05:39:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	332977	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	534620	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	471310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	192730	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	196540	45.52	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.04%		
35) Dibromofluoromethane	5.46	113	157537	46.94	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.88%		
50) Toluene-d8	8.70	98	634972	48.69	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.38%		
62) 4-Bromofluorobenzene	11.12	95	213844	45.58	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.16%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	13648	8.174	ug/l 94
18) Methyl Acetate	2.75	43	3770	1.544	ug/l 99
20) Methylene Chloride	2.83	84	9595	2.299	ug/l 99
43) Isopropyl Acetate	6.42	43	62025	7.842	ug/l 100
44) Trichloroethene	7.18	130	8472	2.100	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019570.D
Acq On : 23 Nov 2020 23:39
Operator : JC/MD
Sample : PB133157TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157TB

Quant Time: Nov 24 05:39:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

