

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019571.D
 Acq On : 24 Nov 2020 00:02
 Operator : JC/MD
 Sample : PB133157ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157ZHE#01

Quant Time: Nov 24 05:41:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	326974	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	537313	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	462167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	197920	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	196987	46.46	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.92%
35) Dibromofluoromethane	5.46	113	158794	47.08	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.16%
50) Toluene-d8	8.70	98	625187	47.70	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.40%
62) 4-Bromofluorobenzene	11.12	95	215447	45.69	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.38%

Target Compounds

					Qvalue
16) Acetone	2.42	43	11976	7.305	ug/l 98
18) Methyl Acetate	2.75	43	4664	1.780	ug/l 98
20) Methylene Chloride	2.83	84	6920	1.689	ug/l 90
43) Isopropyl Acetate	6.42	43	79290	9.974	ug/l 99
44) Trichloroethene	7.18	130	6847	1.689	ug/l 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019571.D
Acq On : 24 Nov 2020 00:02
Operator : JC/MD
Sample : PB133157ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157ZHE#01

Quant Time: Nov 24 05:41:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

