

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019572.D
 Acq On : 24 Nov 2020 00:25
 Operator : JC/MD
 Sample : PB133157ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157ZHE#02

Quant Time: Nov 24 05:46:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	328306	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	532850	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	470878	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199607	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	197419	46.38	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.76%		
35) Dibromofluoromethane	5.46	113	158811	47.48	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.96%		
50) Toluene-d8	8.70	98	631028	48.55	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.10%		
62) 4-Bromofluorobenzene	11.12	95	216577	46.31	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.62%		

Target Compounds

					Qvalue
16) Acetone	2.43	43	11189	6.797	ug/l 97
18) Methyl Acetate	2.75	43	4931	1.841	ug/l 93
20) Methylene Chloride	2.83	84	6673	1.622	ug/l 94
43) Isopropyl Acetate	6.41	43	80709	10.238	ug/l 100
44) Trichloroethene	7.18	130	5948	1.479	ug/l 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019572.D
Acq On : 24 Nov 2020 00:25
Operator : JC/MD
Sample : PB133157ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157ZHE#02

Quant Time: Nov 24 05:46:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

