

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019574.D
 Acq On : 24 Nov 2020 01:12
 Operator : JC/MD
 Sample : PB133157ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157ZHE#04

Quant Time: Nov 24 05:49:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	319558	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	522030	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	464864	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	196426	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	192402	46.44	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.88%	
35) Dibromofluoromethane		5.46	113	152606	46.57	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.14%	
50) Toluene-d8		8.70	98	618246	48.55	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.10%	
62) 4-Bromofluorobenzene		11.12	95	217409	47.46	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	11561	7.215	ug/l	93
18) Methyl Acetate	2.76	43	5275	1.960	ug/l	100
20) Methylene Chloride	2.83	84	6245	1.559	ug/l	99
43) Isopropyl Acetate	6.41	43	78243	10.131	ug/l	99
44) Trichloroethene	7.18	130	4513	1.146	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019574.D
Acq On : 24 Nov 2020 01:12
Operator : JC/MD
Sample : PB133157ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157ZHE#04

Quant Time: Nov 24 05:49:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

