

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019577.D
 Acq On : 24 Nov 2020 02:22
 Operator : JC/MD
 Sample : PB133157ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157ZHE#07

Quant Time: Nov 24 05:56:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	339609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	553744	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	471333	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	197592	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	207532	47.13	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.26%	
35) Dibromofluoromethane		5.46	113	163657	47.08	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.16%	
50) Toluene-d8		8.70	98	649856	48.11	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.22%	
62) 4-Bromofluorobenzene		11.12	95	218220	44.91	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	13191	7.746	ug/l	99
18) Methyl Acetate	2.76	43	4863	1.784	ug/l	96
20) Methylene Chloride	2.83	84	10501	2.467	ug/l	95
43) Isopropyl Acetate	6.41	43	61762	7.539	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019577.D
Acq On : 24 Nov 2020 02:22
Operator : JC/MD
Sample : PB133157ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157ZHE#07

Quant Time: Nov 24 05:56:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

