

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019578.D
 Acq On : 24 Nov 2020 02:45
 Operator : JC/MD
 Sample : PB133157ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157ZHE#08

Quant Time: Nov 24 05:57:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	335204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	540160	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	481960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199966	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	200770	46.19	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.38%	
35) Dibromofluoromethane		5.46	113	161479	47.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.26%	
50) Toluene-d8		8.70	98	653502	49.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.20%	
62) 4-Bromofluorobenzene		11.12	95	218834	46.16	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	13402	7.974	ug/l	98
18) Methyl Acetate	2.76	43	5012	1.835	ug/l	99
20) Methylene Chloride	2.83	84	9950	2.369	ug/l	97
43) Isopropyl Acetate	6.42	43	59087	7.394	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019578.D
Acq On : 24 Nov 2020 02:45
Operator : JC/MD
Sample : PB133157ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157ZHE#08

Quant Time: Nov 24 05:57:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

