

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019580.D
 Acq On : 24 Nov 2020 03:32
 Operator : JC/MD
 Sample : PB133157ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133157ZHE#10

Quant Time: Nov 24 06:02:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	326068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	531854	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	452840	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	187217	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	197337	46.68	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.36%		
35) Dibromofluoromethane	5.46	113	158390	47.44	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.88%		
50) Toluene-d8	8.70	98	618540	47.68	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.36%		
62) 4-Bromofluorobenzene	11.12	95	205944	44.12	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.24%		

Target Compounds

					Qvalue
16) Acetone	2.43	43	12925	7.905	ug/l 99
18) Methyl Acetate	2.75	43	5022	1.871	ug/l 95
20) Methylene Chloride	2.83	84	9474	2.318	ug/l 99
43) Isopropyl Acetate	6.42	43	59894	7.612	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019580.D
Acq On : 24 Nov 2020 03:32
Operator : JC/MD
Sample : PB133157ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133157ZHE#10

Quant Time: Nov 24 06:02:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

