

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019582.D
 Acq On : 24 Nov 2020 04:18
 Operator : JC/MD
 Sample : L4848-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR5-1-TOP

Quant Time: Nov 24 06:07:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	304889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	503080	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	458835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	186007	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	190269	48.13	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.26%		
35) Dibromofluoromethane	5.46	113	150585	47.69	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.38%		
50) Toluene-d8	8.70	98	611030	49.79	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.58%		
62) 4-Bromofluorobenzene	11.12	95	209708	47.50	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.00%		

Target Compounds					Qvalue
16) Acetone	2.42	43	26274	17.187 ug/l	99
18) Methyl Acetate	2.75	43	2906	1.400 ug/l	99
20) Methylene Chloride	2.83	84	52355	13.702 ug/l	96
43) Isopropyl Acetate	6.42	43	23686	3.182 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019582.D
Acq On : 24 Nov 2020 04:18
Operator : JC/MD
Sample : L4848-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR5-1-TOP

Quant Time: Nov 24 06:07:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

