

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019583.D
 Acq On : 24 Nov 2020 04:42
 Operator : JC/MD
 Sample : L4848-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR5-1-BOTTOM

Quant Time: Nov 24 06:10:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	326889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	542425	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	464433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	188997	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	200240	47.24	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.48%
35) Dibromofluoromethane	5.46	113	158754	46.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.26%
50) Toluene-d8	8.70	98	633313	47.87	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.74%
62) 4-Bromofluorobenzene	11.12	95	211290	44.39	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.78%

Target Compounds

					Qvalue
16) Acetone	2.43	43	23027	14.049	ug/l 99
18) Methyl Acetate	2.76	43	3398	1.469	ug/l 96
20) Methylene Chloride	2.83	84	36530	8.917	ug/l 96
43) Isopropyl Acetate	6.41	43	21105	2.630	ug/l 98
44) Trichloroethene	7.18	130	4117	1.006	ug/l 96
95) Naphthalene	13.82	128	6107	2.597	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019583.D
Acq On : 24 Nov 2020 04:42
Operator : JC/MD
Sample : L4848-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR5-1-BOTTOM

Quant Time: Nov 24 06:10:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

