

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019584.D
 Acq On : 24 Nov 2020 05:05
 Operator : JC/MD
 Sample : L4848-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR5-2-TOP

Quant Time: Nov 24 07:04:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	331678	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	535805	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	453141	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	193092	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	197987	46.04	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.08%		
35) Dibromofluoromethane	5.46	113	156901	46.65	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.30%		
50) Toluene-d8	8.70	98	623349	47.70	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.40%		
62) 4-Bromofluorobenzene	11.12	95	212167	45.12	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.24%		

Target Compounds

					Qvalue
16) Acetone	2.43	43	25470	15.315	ug/l 99
18) Methyl Acetate	2.75	43	2920	1.342	ug/l 98
20) Methylene Chloride	2.83	84	41102	9.888	ug/l 99
43) Isopropyl Acetate	6.42	43	21712	2.739	ug/l 98
95) Naphthalene	13.82	128	3201	2.393	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019584.D
Acq On : 24 Nov 2020 05:05
Operator : JC/MD
Sample : L4848-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR5-2-TOP

Quant Time: Nov 24 07:04:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

