

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019586.D
 Acq On : 24 Nov 2020 05:52
 Operator : JC/MD
 Sample : L4848-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR5-3-TOP

Quant Time: Nov 24 07:08:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	316820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	515332	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	445152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	171744	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	192980	46.98	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.96%		
35) Dibromofluoromethane	5.46	113	152037	47.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.00%		
50) Toluene-d8	8.70	98	616738	49.06	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.12%		
62) 4-Bromofluorobenzene	11.12	95	199168	44.04	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.08%		

Target Compounds					Qvalue
16) Acetone	2.42	43	30474	19.183	ug/l 99
18) Methyl Acetate	2.75	43	2234	1.201	ug/l 98
20) Methylene Chloride	2.83	84	46676	11.756	ug/l 98
43) Isopropyl Acetate	6.42	43	18935	2.484	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019586.D
Acq On : 24 Nov 2020 05:52
Operator : JC/MD
Sample : L4848-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
TR5-3-TOP

Quant Time: Nov 24 07:08:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

