

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019588.D
 Acq On : 24 Nov 2020 06:38
 Operator : JC/MD
 Sample : L4848-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR5-4-TOP

Quant Time: Nov 24 07:13:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	317433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	517524	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	455702	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	193421	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	195012	47.38	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.76%		
35) Dibromofluoromethane	5.46	113	153351	47.21	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.42%		
50) Toluene-d8	8.70	98	617596	48.92	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.84%		
62) 4-Bromofluorobenzene	11.12	95	209578	46.15	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.30%		

Target Compounds					Qvalue
16) Acetone	2.42	43	32478	20.405	ug/l 95
18) Methyl Acetate	2.75	43	2141	1.176	ug/l 97
20) Methylene Chloride	2.83	84	45172	11.355	ug/l 96
43) Isopropyl Acetate	6.41	43	15765	2.059	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019588.D
Acq On : 24 Nov 2020 06:38
Operator : JC/MD
Sample : L4848-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR5-4-TOP

Quant Time: Nov 24 07:13:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

