

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019590.D
 Acq On : 24 Nov 2020 07:25
 Operator : JC/MD
 Sample : L4858-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PHC-340EM-012

Quant Time: Nov 24 08:13:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	327989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	535142	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	462972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199235	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	200971	47.26	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.52%	
35) Dibromofluoromethane		5.46	113	159486	47.48	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.96%	
50) Toluene-d8		8.70	98	636659	48.77	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.54%	
62) 4-Bromofluorobenzene		11.12	95	213619	45.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	28240	17.172	ug/l	95
18) Methyl Acetate	2.75	43	31164	8.265	ug/l	99
20) Methylene Chloride	2.83	84	47443	11.542	ug/l	97
43) Isopropyl Acetate	6.42	43	26360	3.329	ug/l	99
95) Naphthalene	13.82	128	949	2.240	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019590.D
Acq On : 24 Nov 2020 07:25
Operator : JC/MD
Sample : L4858-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PHC-340EM-012

Quant Time: Nov 24 08:13:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

