

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112320\
 Data File : VX019591.D
 Acq On : 24 Nov 2020 07:48
 Operator : JC/MD
 Sample : L4858-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PHC-340EM-013

Quant Time: Nov 24 08:15:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	325515	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	535690	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	454038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	192137	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	201769	47.81	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.62%	
35) Dibromofluoromethane		5.46	113	158575	47.16	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.32%	
50) Toluene-d8		8.70	98	625785	47.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.78%	
62) 4-Bromofluorobenzene		11.12	95	209205	44.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	48641	29.801	ug/l	99
18) Methyl Acetate	2.75	43	480658	119.217	ug/l	100
20) Methylene Chloride	2.83	84	48939	11.996	ug/l	99
43) Isopropyl Acetate	6.42	43	29017	3.661	ug/l	99
95) Naphthalene	13.82	128	7204	2.664	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112320\
Data File : VX019591.D
Acq On : 24 Nov 2020 07:48
Operator : JC/MD
Sample : L4858-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PHC-340EM-013

Quant Time: Nov 24 08:15:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

