

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
 Data File : VX025268.D
 Acq On : 23 Nov 2021 12:47
 Operator : JC/MD
 Sample : M4801-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20211119

Quant Time: Nov 23 15:20:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196740	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64333	55.585	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.180%
35) Dibromofluoromethane	5.391	113	62602	51.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.720%
50) Toluene-d8	8.653	98	232245	51.170	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.340%
62) 4-Bromofluorobenzene	11.079	95	86937	50.632	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.260%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	7833	19.102	ug/l	98
16) Acetone	2.380	43	9015	14.018	ug/l	99
25) 2-Butanone	4.562	43	2704	2.908	ug/l	99
31) Cyclohexane	5.477	56	3901	2.181	ug/l	89
68) m/p-Xylenes	10.305	106	1086	0.398	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112321\
Data File : VX025268.D
Acq On : 23 Nov 2021 12:47
Operator : JC/MD
Sample : M4801-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20211119

Quant Time: Nov 23 15:20:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

